



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/649,677	08/28/2003	Andreas Peter Abel	2003_1239	5670

513 7590 08/02/2006

WENDEROTH, LIND & PONACK, L.L.P.
2033 K STREET N. W.
SUITE 800
WASHINGTON, DC 20006-1021

EXAMINER

LEE, SHUN K

ART UNIT	PAPER NUMBER
----------	--------------

2884

DATE MAILED: 08/02/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

5/

Office Action Summary

Application No.

10/649,677

Applicant(s)

ABEL ET AL.

Examiner

Shun Lee

Art Unit

2884

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 08 May 2006.
 2a) ☒ This action is FINAL. 2b) ☐ This action is non-final.
 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,2,4-10,12-27,29-40 and 56-59 is/are pending in the application.
 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
 5) ☐ Claim(s) _____ is/are allowed.
 6) ☒ Claim(s) 1,2,4-10,12-27,29-40 and 56-59 is/are rejected.
 7) ☐ Claim(s) _____ is/are objected to.
 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
 10) ☒ The drawing(s) filed on 28 August 2003 and 06 January 2005 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.

Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).

Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).

- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☒ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☒ Certified copies of the priority documents have been received in Application No. 10/000,957.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
 Paper No(s)/Mail Date _____
 4) ☐ Interview Summary (PTO-413)
 Paper No(s)/Mail Date. _____
 5) ☐ Notice of Informal Patent Application (PTO-152)
 6) ☐ Other: _____

DETAILED ACTION

Specification

1. The lengthy specification has not been checked to the extent necessary to determine the presence of all possible minor errors. Applicant's cooperation is requested in correcting any errors of which applicant may become aware in the specification.

Claim Rejections - 35 USC § 103

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

4. Claims 1, 2, 4-10, 12-17, 19-27, 29-40, and 56-59 are rejected under 35 U.S.C. 103(a) as being unpatentable over Neuschäfer *et al.* (WO 96/35940) in view of Fattinger (US 5,455,178) and Rudigier *et al.* (US 5,738,825).

In regard to claims **1**, **23**, **36**, **56**, **58**, and **59**, Neuschäfer *et al.* disclose (Figs. 1c and 6) an analytical system for the determination of one or more analytes in at least one sample by luminescence detection, said analytical system comprising:

(a) at least one excitation light source (13) operable to emit excitation light;

(b) a sensor platform (8; for a simultaneous determination of at least one luminescence from a plurality of measurement areas) comprising:

(b1) a plurality of laterally separated measurement areas (1), wherein a density of said plurality of laterally separated measurement areas is 0.4 to 40,000 (e.g., at least 16) measurement areas per square centimeter (*i.e.*, each measurement area is 5 μm to 5 mm width by 0.5 to 50 mm length or equivalently $2.5 \times 10^{-5} \text{ cm}^2$ to 2.5 cm^2 ; last paragraph on pg. 12 to fourth paragraph on pg. 13);

(b2) an optical film waveguide (first two paragraphs on pg. 1) comprising a waveguiding layer (*i.e.*, a first optically transparent layer), a substrate (*i.e.*, a second optically transparent layer) having a lower refractive index than said first optically transparent layer, said first optically transparent layer being on said second optically transparent layer, and a grating structure (3 and/or 3') being operable to incouple the excitation light to said plurality of laterally separated measurement areas (1), wherein said plurality of laterally separated measurement areas (1) are located on said first optically transparent layer, said grating structure (3 and/or 3') is continuously modulated (third paragraph on pg. 8; third paragraph on pg. 17) over a complete of said plurality of

- laterally separated measurement areas (1), and said grating structure (3 and/or 3') is operable to prevent a cross-talk (fourth paragraph on pg. 7) of luminescence generated in any one measurement area of said plurality of laterally separated measurement areas (1), and coupled back into said first optically transparent layer to any other measurement area of said plurality of laterally separated measurement areas; and
- (b3) at least one recognition element (second paragraph on pg. 18) immobilized in said plurality of laterally separated measurement areas (1), said at least one recognition element being operable to assist in a qualitative or quantitative determination of the one or more analytes in the at least one sample in contact with said plurality of laterally separated measurement areas (1);
- (c) at least one detector (14) operable to collect light emanating from one or more of said plurality of laterally separated measurement areas (1) on said sensor platform (8); and
- (d) supply means (10, 11; second paragraph on pg. 26) for supplying the at least one sample in contact with said plurality of laterally separated measurement areas (1) on said sensor platform (8).

While Neuschäfer *et al.* also disclose (fourth paragraph on pg. 17; fifth paragraph on pg. 28) that the grating structure comprises of plurality of coupling-in and coupling-out gratings having different grating constants (*i.e.*, periodicities) for the analysis of samples of different luminescence and that excitation of luminescence requires light sources of identical or different wavelengths, the system of Neuschäfer *et al.* lacks that the grating

Art Unit: 2884

structure is a superposition of a plurality of grating structures of different periodicities for the incoupling of excitation light of different wavelengths, wherein the grating structure has a laterally varying periodicity either in parallel or perpendicular to a direction of propagation of the incoupled excitation light in the first optically transparent layer.

However, gratings for the coupling of laser beams sources are well known in the art.

For example, Fattinger teaches (Figs. 1-3) that gratings can be unidiffractive or multidiffractive (*i.e.*, superposition of two or more gratings) for coupling of light from different directions. Further, Rudigier *et al.* teach (Fig. 5) that gratings for light coupling can have a laterally varying periodicity in parallel or perpendicular to the direction of propagation of the in-coupled light. Therefore it would have been obvious to one having ordinary skill in the art at the time of the invention to provide a multidiffractive grating having a laterally varying periodicity grating as the grating structure in the system of Neuschäfer *et al.*, in order to analyze samples of different luminescence which require luminescence excitation from different wavelength light sources located at different directions from the grating structure.

In regard to claim 2 (which is dependent on claim 1) and claim 57 (which is dependent on claim 56), Neuschäfer *et al.* also disclose (Figs. 2-5) that said plurality of laterally separated measurement areas are split into at least two laterally separated segments, each of said laterally separated segments comprising at least two of said plurality of laterally separated measurement areas.

In regard to claim 4 which is dependent on claim 1, Neuschäfer *et al.* also disclose (fourth and fifth paragraphs on pg. 15) an intermediate layer (*i.e.*, a third

Art Unit: 2884

optically transparent layer) in contact with said first optically transparent layer, wherein said third optically transparent layer has a lower refractive index than said first optically transparent layer and said third optically transparent layer has a thickness of $\gg 100$ nm and not thicker than 0.01 mm (*i.e.*, 5 nm - 10,000 nm) and is located between said first optically transparent layer and said second optically transparent layer.

In regard to claim 5 which is dependent on claim 1, Neuschäfer *et al.* also disclose (fifth through seventh paragraphs on pg. 19; last paragraph on pg. 20) an adhesion-promoting layer with a thickness of less than 50 nm (*i.e.*, less than 200 nm) deposited on said first optically transparent layer, said adhesion-promoting layer operable to immobilize one of biological elements, biochemical elements and synthetic recognition elements, wherein said adhesion-promoting layer comprises chemical compounds from a group consisting of silanes, epoxides, and self-organized functionalized monolayers.

In regard to claim 6 which is dependent on claim 1, Neuschäfer *et al.* also disclose (eighth paragraph on pg. 19) that said plurality of laterally separated measurement areas are generated by laterally selective deposition of at least one of biological elements, biochemical elements and synthetic recognition elements on said sensor platform, by one of jet spotting, mechanical spotting, micro contact printing, and fluidic contacting said plurality of laterally separated measurement areas with said at least one of biological elements, biochemical elements and synthetic recognition elements supplied in parallel or by crossed micro channels, upon application of one of pressure differences, electric potentials and electromagnetic potentials.

In regard to claim **7** which is dependent on claim 6, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 21) that said at least one of biological elements, biochemical elements and synthetic recognition elements, components of a group consisting of nucleic acids and nucleic acid analogues, antibodies, aptamers, membrane-bound and isolated receptors, ligands of the membrane-bound and isolated receptors, antigens for antibodies, histidine-tag components, and molecular imprints hosted in cavities generated by chemical synthesis, are deposited as whole cells or cell fragments.

In regard to claim **8** which is dependent on claim 6, Neuschäfer *et al.* also disclose (eighth paragraph on pg. 19) compounds, which are chemically neutral towards the at least one analyte, deposited between said plurality of laterally separated measurement areas in order to minimize nonspecific binding or adsorption.

In regard to claim **9** which is dependent on claim 1, Neuschäfer *et al.* also disclose (last paragraph on pg. 12 to fourth paragraph on pg. 13) that 2 to 1000 (*i.e.*, up to 100,000) of 5 μm to 5 mm width by 0.5 to 50 mm length measurement areas are provided in a 2-dimensional arrangement (see Figs. 1-5) and thus a single measurement area has an area of 0.0025 mm^2 to 250 mm^2 (*i.e.*, 0.001 mm^2 - 6 mm^2).

In regard to claim **10** which is dependent on claim 1, Neuschäfer *et al.* also disclose (last three paragraphs on pg. 16 and first seven paragraphs on pg. 17) that said grating structure is one of a diffractive grating with a uniform period and a multidiffractive grating.

In regard to claim **12** which is dependent on claim 1, Neuschäfer *et al.* also disclose (second paragraph on pg. 15) that the substrate (*i.e.*, said second optically transparent layer) comprises one of quartz, glass or transparent thermoplastic.

In regard to claim **13** which is dependent on claim 1, Neuschäfer *et al.* also disclose (first paragraph on pg. 16) that said first optically transparent layer has a refractive index that is higher than 2.

In regard to claim **14** which is dependent on claim 1, Neuschäfer *et al.* also disclose (second paragraph on pg. 16) that said first optically transparent layer comprises one of TiO₂, ZnO, Nb₂O₅, Ta₂O₅, HfO₂, or ZrO₂.

In regard to claim **15** which is dependent on claim 1, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 16) that said first optically transparent layer has a thickness of between 40 and 1,000 nm (*i.e.*, between 40 and 300 nm).

In regard to claim **16** which is dependent on claim 1, Neuschäfer *et al.* also disclose (fifth paragraph and also the last paragraph on pg. 16) that said grating structure has a period of 200 nm - 1000 nm and a modulation depth of 3 nm - 60 nm (*i.e.*, 3 nm - 100 nm).

In regard to claim **17** which is dependent on claim 1, Neuschäfer *et al.* also disclose (third and fourth paragraphs on pg. 4) that coupling efficiency of the grating coupler depends on the grating coupler modulation depth (with reference to Proc. SPIE 2068:313-325) and that coupling efficiency should be adequate so as to be able to measure luminescence. Thus inherent in an adequate coupling efficiency (that depends on the grating coupler modulation depth) is a coupling efficiency and is less than 100%

(*i.e.*, incoupling and outcoupling of excitation light and/or backcoupled luminescence light is incomplete). Neuschäfer *et al.* further disclose (seventh paragraph on pg. 17) that the guided radiation intensity falls and the length at which the guided radiation intensity has fallen to I_0/e is $X_{1/e}$. Thus it is inherent in guided radiation which has fallen to I_0/e at a distance of $X_{1/e}$ that the extent of the losses along the guided radiation propagation direction (*i.e.*, distance $X_{1/e}$) determines a gradient of the guided radiation intensity (*i.e.*, at least one of intensity of guided excitation light and generated luminescence light) in the waveguiding layer (*i.e.*, first optically transparent layer) within at least one measurement area of said plurality of laterally separated measurement areas and across several measurement areas of said plurality of laterally separated measurement areas.

In regard to claim **19** which is dependent on claim 1, Neuschäfer *et al.* also disclose (seventh paragraph on pg. 17) that the guided radiation intensity falls and the length at which the guided radiation intensity has fallen to I_0/e is $X_{1/e}$. Thus it is inherent in a guided radiation which has fallen to I_0/e at a distance of $X_{1/e}$ that the extent of the losses along the guided radiation propagation direction (*i.e.*, distance $X_{1/e}$) determines a gradient of the guided radiation intensity (*i.e.*, at least one of intensity of guided excitation light and generated luminescence light) in the waveguiding layer (*i.e.*, first optically transparent layer) within at least one of measurement area of said plurality of laterally separated measurement areas and across several measurement areas said plurality of laterally separated measurement areas.

In regard to claim **20** which is dependent on claim 16, Neuschäfer *et al.* also disclose (seventh paragraph on pg. 16) that a ratio of the modulation depth to a thickness of said first optically transparent layer is equal or smaller than 0.2.

In regard to claim **21** which is dependent on claim 1, Neuschäfer *et al.* also disclose (eighth and ninth paragraphs on pg. 16) that said grating structure is one of a relief grating with a rectangular, triangular or semi-circular profile and a phase or volume grating with a periodic modulation of a refractive index in said first optically transparent layer which is essentially planar.

In regard to claim **22** which is dependent on claim 1, Neuschäfer *et al.* also disclose (last paragraph on pg. 39) one of optically recognizable marks and mechanically recognizable marks operable to simplify adjustments in an optical system, or for connection to sample compartments as part of an analytical system.

In regard to claim **24** which is dependent on claim 23, Neuschäfer *et al.* also disclose (last two paragraphs on pg. 7) that said at least one excitation light source launches the excitation light to said plurality of laterally separated measurement areas in an arrangement of direct or transmission illumination.

In regard to claim **25** which is dependent on claim 23, Neuschäfer *et al.* also disclose (second paragraph on pg. 29) that said at least one detector also collects luminescence light outcoupled by said grating structure.

In regard to claim **26** which is dependent on claim 23, Neuschäfer *et al.* also disclose (last paragraph on pg. 16; last two paragraphs on pg. 23; third paragraph on pg. 28) that the excitation light emitted from said at least one excitation light source is

coherent and is launched to said plurality of laterally separated measurement areas at the first diffraction order (*i.e.*, a resonance angle) for coupling into the waveguiding layer (*i.e.*, first optically transparent layer).

In regard to claim **27** which is dependent on claim 23, Neuschäfer *et al.* also disclose (third paragraph on pg. 28) that said at least one excitation light source is a plurality of coherent light sources of one of similar and different emission wavelengths.

In regard to claim **29** which is dependent on claim 23, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 26) that said at least one detector is a laterally resolving detector from a group consisting of CCD cameras, CCD chips, photodiode arrays, avalanche diode allays, multichannel plates and multichannel photomultipliers.

In regard to claim **30** which is dependent on claim 23, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 26) at least one optical component being located between at least one of said at least one excitation light source and said sensor platform and said sensor platform and said at least one detector, said at least one optical component comprising at least one of: a lens or a lens system operable to shape at least one of the excitation light and the one or more luminescences; a planar mirror or a curved mirror for deviation of at least one of the excitation light and the one or more luminescences; a prism for deviation of at least one of the excitation light and the one or more luminescences; a dichroic mirror for the spectrally selective deviation of parts of at least one of the excitation light and the one or more luminescences; a neutral density filter for regulation of light intensity of at least one of the excitation light and the one or more luminescences; an optical filter or a monochromator for spectrally selective

Art Unit: 2884

transmission of parts of at least one of the excitation light and the one or more luminescences; and a polarization selective element for selection of discrete polarization directions of at least one of the excitation light and the one or more luminescences.

In regard to claim **31** which is dependent on claim 23, Neuschäfer *et al.* also disclose (last paragraph on pg. 27) that said at least one excitation light source launches the excitation light in pulses with a duration of 1 ps to 100 s (*i.e.*, 1 fs to 10 min) and emission light from said plurality of laterally separated measurement areas is measured time-resolved.

In regard to claims **32** and **33** which are dependent on claim 23, Neuschäfer *et al.* also disclose (third paragraph on pg. 27) that said at least one detector measures light signals from at least one of the excitation light at a location of said at least one excitation light source, the excitation light after expansion, the excitation light after being multiplexed into individual beams, scattered excitation light from a location of one or more measurement areas of said plurality of laterally separated measurement areas, and the excitation light outcoupled by said grating structure beside said plurality of measurement areas for referencing purposes as a reference signal and that said one or more of said plurality of measurement areas for determination of the emission light and for the determination of the reference signal are the same.

In regard to claim **34** which is dependent on claim 23, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 28) that said at least one excitation light source and

said at least one detector respectively launch the excitation light and detect the emission light from said plurality of laterally separated measurement areas sequentially for one or more measurement areas of said plurality of laterally separated measurement areas.

In regard to claim **35** which is dependent on claim 34, Neuschäfer *et al.* also disclose (fourth paragraph on pg. 28) that said sensor platform is operable to be moved between the sequential excitation and detection.

In regard to claims **37-39** which are dependent on claim 36, Neuschäfer *et al.* also disclose (first four paragraphs on pg. 14; second paragraph on pg. 26) at least one sample compartment (12) which is at least in the area of at least one laterally separated measurement area of said plurality of laterally separated measurement areas (1), wherein said at least one sample compartment (12) has a volume of 0.1 nL - 100 μ L (e.g., 28 mm by 6 mm by 0.2 mm or about 37 μ L), said at least one sample compartment (12) has an opening for supplying and/or removing samples or other reagents at a side opposite to the waveguiding layer (*i.e.*, first optically transparent layer), and said at least one sample compartment (12) being otherwise closed, and wherein the supplying and removing of the samples is performed in a closed flow through system.

In regard to claim **40** which is dependent on claim 37, Neuschäfer *et al.* also disclose (fifth paragraph on pg. 22; second and third paragraphs on pg. 24) that multi-step assays require multiple solutions. Thus the at least one sample compartment (see also fifth paragraph on pg. 39) is capable of accepting reagents (e.g., a luminescence-

Art Unit: 2884

labeled antibody) such that the reagents can be wetted during an assay for the determination of the one or more analytes and in contact with said at least one laterally separated measurement area.

5. Claim 18 is rejected under 35 U.S.C. 103(a) as being unpatentable over Neuschäfer *et al.* (WO 96/35940) in view of Fattinger (US 5,455,178) and Rudigier *et al.* (US 5,738,825) as applied to claim 17 above, and further in view of Sunagawa (US 5,101,459).

In regard to claim **18** which is dependent on claim 17, the system of Neuschäfer *et al.* lacks that said grating structure has a laterally varying grating depth in parallel with the direction of propagation of the incoupled excitation light. However, coupling of laser beams sources and waveguides is well known in the art. For example, Sunagawa teaches (column 3, lines 22-68) it is known in the art that the grating structure has a laterally varying grating depth in parallel to the direction of propagation of the incoupled excitation light in order to enhance coupling efficiency of laser beams sources and waveguides. Therefore it would have been obvious to one having ordinary skill in the art at the time of the invention to provide a laterally varying grating depth grating structure in the system of Neuschäfer *et al.*, in order to enhance coupling efficiency between laser beams sources and the waveguiding layer.

Double Patenting

6. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA

Art Unit: 2884

1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

7. Claim 1 is provisionally rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claim 21 of US 2003/0138208 A1 (copending Application No. 10/275,380). Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims in the application define an invention that is merely an obvious variation of the invention claimed in copending Application No. 10/275,380.

This is a provisional obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

Response to Arguments

8. Applicant's arguments filed 8 May 2006 have been fully considered but they are not persuasive.

Applicant argues (first paragraph on pg. 24 to second paragraph on pg. 25 of remarks filed 8 May 2006) that the combination of Neuschäfer *et al.* with Fattinger does not result in the invention as presently claimed, with the advantageous uses of the claimed invention as disclosed in the specification since Fattinger does not discuss the use of luminophores of different excitation wavelengths and similar or different emission wavelengths (depending on the experiment to be performed), which would be

Art Unit: 2884

advantageous in the field of biology (a field not discussed by Fattinger). In response to applicant's argument, the fact that applicant has recognized another advantage which would flow naturally from following the suggestion of the prior art cannot be the basis for patentability when the differences would otherwise be obvious. See *Ex parte Obiaya*, 227 USPQ 58, 60 (Bd. Pat. App. & Inter. 1985). In addition, Neuschäfer *et al.* state (fourth paragraph on pg. 17) that "For the analysis of samples of different luminescences, it may be advantageous for all or some of the coupling-in and coupling-out gratings to have different grating constants" and (fifth paragraph on pg. 28) that "Another preferred method consists in exciting the luminescences with various laser light sources of identical or different wavelengths". Thus Neuschäfer *et al.* expressly teach the use of luminophores of different excitation wavelengths and similar or different emission wavelengths. Further, Fattinger state (column 1, line 10 to column 4, line 56) that "The invention relates to a microoptical sensor for detecting chemical substances, this sensor having a planar optical waveguide which comprises a solid substrate having layered on its top surface a waveguiding layer and a diffraction grating, situated on or in the immediate vicinity of the waveguiding layer, for coupling and decoupling coherent light applied thereto. ... Many of the conventional test methods used in biomedical diagnosis ... In conventional tests, detection of the molecules bound to the chemosensitive coating is usually effected indirectly by means of a second dissolved reaction partner labelled with a radio isotope, a fluorophore or an enzyme ... ". Thus Fattinger teaches gratings for optical measurements in applications such as biomedical diagnosis (e.g., detection of the molecules using fluorophore labels). Therefore it would

Art Unit: 2884

have been obvious to one having ordinary skill in the art at the time of the invention to provide a multidiffractive grating having a laterally varying periodicity grating as the grating structure in the system of Neuschäfer *et al.*, in order to analyze samples of different luminescence which require luminescence excitation from different wavelength light sources located at different directions from the grating structure.

In response to applicant's argument (third paragraph on pg. 25 to second paragraph on pg. 26 of remarks filed 8 May 2006) that the references fail to show certain features of applicant's invention, it is noted that the features upon which applicant relies (*i.e.*, using luminescences in a polarization-selective way) are not recited in the rejected claim(s). Although the claims are interpreted in light of the specification, limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993).

Applicant's arguments (fourth and fifth paragraphs on pg. 26 of remarks filed 8 May 2006) with respect to claim 18 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

9. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within


TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

10. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Shun Lee whose telephone number is (571) 272-2439. The examiner can normally be reached on Monday-Thursday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David Porta can be reached on (571) 272-2444. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

SL



DAVID PORTA
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 2800